



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

AppMet, Inc.
7308 Peppermill Parkway
North Charleston, SC 29418

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 06 December 2024

Certificate Number: AC-1358



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AppMet, Inc.
 7308 Peppermill Parkway
 North Charleston, SC 29418
 Melinda Bigg 843-767-5664
 mbigg@appmet.com

CALIBRATION

Valid to: **December 6, 2024**

Certificate Number: **AC-1358**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source ¹	Up to 330 mV 330 mV to 3.3 V (3.3 to 33) V (33 to 330) V 330 V to 1 kV	23 μ V/V + 1.2 μ V 14 μ V/V + 2.4 μ V 13 μ V/V + 23 μ V 21 μ V/V + 0.17 mV 21 μ V/V + 1.7 mV	Fluke 5520A Multifunction Calibrator
AC Voltage – Source ¹	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz 330 mV to 3.3 V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.91 mV/V + 6.9 μ V 0.18 mV/V + 7 μ V 0.25 mV/V + 6.9 μ V 1.2 mV/V + 6.9 μ V 4 mV/V + 14 μ V 9.2 mV/V + 58 μ V 0.36 mV/V + 9 μ V 0.17 mV/V + 9.6 μ V 0.19 mV/V + 9 μ V 0.41 V/V + 9.8 μ V 0.93 V/V + 37 μ V 2.3 mV/V + 81 μ V 0.34 mV/V + 61 μ V 0.16 mV/V + 0.1 mV 0.22 mV/V + 95 μ V 0.33 mV/V + 65 μ V 0.79 mV/V + 0.15 mV 2.7 mV/V + 0.72 mV	Fluke 5520A Multifunction Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (33 to 330) V 45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz 330 V to 1.02 kV 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.37 mV/V + 0.64 mV 0.18 mV/V + 0.69 mV 0.28 mV/V + 0.71 mV 0.40 mV/V + 0.77 mV 1 mV/V + 1.9 mV 0.22 mV/V + 2.3 mV 0.23 mV/V + 7.4 mV 0.29 mV/V + 6.6 mV 0.35 mV/V + 7.1 mV 2.2 mV/V + 63 mV 0.36 mV/V + 7.8 mV 0.29 mV/V + 13 mV 0.37 mV/V + 7.9 mV	Fluke 5520A Multifunction Calibrator
DC Current – Source ¹	Up to 330 μ A 330 μ A to 3.3 mA (3.3 to 33) mA (33 to 330) mA 330 mA to 1.1 A (1.1 to 3) A (3 to 11) A (11 to 20.5) A	0.18 mA/A + 23 nA 0.12 mA/A to 41 nA 0.2 mA/A + 0.19 μ A 0.2 mA/A + 19 nA 0.23 mA/A + 52 μ A 0.45 mA/A + 37 μ A 0.58 mA/A + 0.58 mA 1.1 mA/A + 1.8 mA	Fluke 5520A Multifunction Calibrator
AC Current – Source ¹	(29 to 330) μ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 330 μ A to 3.3 mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	2.3 mA/A + 0.12 μ A 1.7 mA/A + 0.11 μ A 1.4 mA/A + 0.1 μ A 3.3 mA/A + 0.17 μ A 6 mA/A + 0.32 μ A 2.3 mA/A + 0.17 μ A 1.7 mA/A + 0.08 μ A 1.1 mA/A + 0.27 μ A 2.3 mA/A + 0.23 μ A 3 mA/A + 1.3 μ A	Fluke 5520A Multifunction Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
AC Current – Source ¹	(3.3 to 33) mA		Fluke 5520A Multifunction Calibrator		
	(10 to 20) Hz	9.3 mA/A + 0.6 μA			
	(20 to 45) Hz	1.1 mA/A + 2.2 μA			
	45 Hz to 1 kHz	0.41 mA/A + 2.6 μA			
	(1 to 5) kHz	0.72 mA/A + 2.9 μA			
	(5 to 10) kHz	2.3 mA/A + 3.5 μA			
	(33 to 330) mA				
	(10 to 20) Hz	2.1 mA/A + 23 μA			
	(20 to 45) Hz	1.1 mA/A + 23 μA			
	45 Hz to 1 kHz	0.46 mA/A + 23 μA			
	(1 to 5) kHz	1.2 mA/A + 58 μA			
	(5 to 10) kHz	2.3 mA/A + 0.12 mA			
	330 mA to 1.1 A				
	(10 to 45) Hz	2.1 mA/A + 0.10 mA			
	45 Hz to 1 kHz	0.42 mA/A + 0.17 mA			
	(1 to 5) kHz	6.9 mA/A + 1.2 mA			
	Resistance – Source ¹	Up to 11 Ω			Fluke 5520A Multifunction Calibrator
		(11 to 33) Ω		37 μΩ/Ω + 1.4 mΩ	
(33 to 110) Ω		35 μΩ/Ω + 1.8 mΩ			
(110 to 330) Ω		58 μΩ/Ω + 0.8 mΩ			
330 Ω to 1.1 kΩ		23 μΩ/Ω + 5.2 mΩ			
(1.1 to 3.3) kΩ		33 μΩ/Ω + 1.8 mΩ			
(3.3 to 11) kΩ		31 μΩ/Ω + 25 mΩ			
(11 to 33) kΩ		33 μΩ/Ω + 18 mΩ			
(33 to 110) kΩ		31 μΩ/Ω + 0.25 Ω			
(110 to 330) kΩ		33 μΩ/Ω + 0.18 Ω			
330 kΩ to 1.1 MΩ		40 μΩ/Ω + 2.1 Ω			
(1.1 to 3.3) MΩ		36 μΩ/Ω + 4.2 Ω			
(3.3 to 11) MΩ		62 μΩ/Ω + 78 Ω			
(11 to 33) MΩ		0.14 mΩ/Ω + 0.33 kΩ			



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source ¹	(11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ	0.29 mΩ/Ω + 3.4 kΩ 0.37 mΩ/Ω + 46 kΩ 3.5 mΩ/Ω + 0.13 MΩ	Fluke 5520A Multifunction Calibrator
Resistance – Source ¹	1GΩ 10GΩ 100GΩ 1TΩ	0.013 GΩ 0.12 GΩ 1.7 GΩ 0.15 TΩ	Fixed Resistors
DC Voltage – Measure ¹	Up to 100 mV 100 mV to 1 V (1 to 10) V (10 to 100) V 100 V to 1 kV	4.4 μV/V + 0.41 μV 3.2 μV/V + 0.82 μV 2.8 μV/V + 3.3 μV 5.6 μV/V + 2.8 μV 1.7 μV/V + 0.15 mV	HP 3458A Precision Multimeter
Capacitance – Source ¹	190 pF to 3.3 nF (3.3 to 11) nF (11 to 110) nF (110 to 330) nF 330 nF to 1.1 μF (1.1 to 3.3) μF (3.3 to 11) μF (11 to 33) μF (33 to 110) μF (110 to 330) μF 330 μA to 1.1 mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	5.9 mF/F + 12 pF 6.5 mF/F + 10 pF 4.8 mF/F + 29 pF 3.6 mF/F + 65 pF 3.7 mF/F + 57 pF 3 mF/F + 1.1 nF 2.8 mF/F + 3.6 nF 3.1 mF/F + 11 nF 4.5 mF/F + 36 nF 5.3 mF/F + 0.11 μF 5.2 mF/F + 0.36 μF 4.5 mF/F + 1.1 μF 5.3 mF/F + 3.5 μF 5.2 mF/F + 12 μF 8.8 mF/F + 30 μF 13 mF/F + 0.11 mF	Fluke 5520A Multifunction Calibrator
AC Voltage – Measure ¹	Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	12 μV/V + 3.4 μV 0.22 mV/V + 1.5 μV 0.35 mV/V + 1.3 μV 1.2 mV/V + 1.3 μV 5.8 mV/V + 1.4 μV 46 mV/V + 2.4 μV 86 μV/V + 4.3 μV 81 μV/V + 2.5 μV 0.17 mV/V + 1.6 μV 0.34 mV/V + 2.8 μV 0.92 mV/V + 2.6 μV 3.5 mV/V + 12 μV	HP 3458A Precision Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	100 mV to 1 V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (1 to 10) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz 100 V to 1 kV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	79 μ V/V + 50 μ V 80 μ V/V + 25 μ V 0.16 mV/V + 23 μ V 0.35 mV/V + 24 μ V 0.92 mV/V + 23 μ V 3.5 mV/V + 0.12 mV 12 mV/V + 0.14 mV 17 mV/V + 0.13 mV 85 μ V/V + 0.46 mV 83 μ V/V + 0.22 mV 0.16 mV/V + 0.23 mV 0.35 mV/V + 0.23 mV 0.92 mV/V + 0.23 mV 3.5 mV/V + 1.2 mV 11 mV/V + 12 mV 17 mV/V + 10 mV 0.23 mV/V + 4.7 mV 0.23 mV/V + 2.3 mV 0.23 mV/V + 2.5 mV 0.23 mV/V + 2.3 mV 0.23 mV/V + 2.3 mV 0.41 mV/V + 2.3 mV 4.6 mV/V + 12 mV 17 mV/V + 12 mV 0.46 mV/V + 46 mV 0.46 mV/V + 23 mV 0.23 mV/V + 23 mV 1.4 mV/V + 23 mV 3.5 mV/V + 23 mV	HP 3458A Precision Multimeter
	DC Current – Measure ¹	100 nA to 1 μ A (1 to 10) μ A (10 to 100) μ A 100 μ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ¹	Up to 100 μ A		HP 3458A Precision Multimeter
	(10 to 20) Hz	4.6 mA/A + 35 nA	
	(20 to 45) Hz	1.7 mA/A + 35 nA	
	(45 to 100) Hz	0.7 mA/A + 35 nA	
	100 Hz to 1 kHz	0.7 mA/A + 35 nA	
	100 μ A to 1 mA		
	(10 to 20) Hz	4.5 mA/A + 0.32 μ A	
	(20 to 45) Hz	1.7 mA/A + 0.23 μ A	
	(45 to 100) Hz	0.7 mA/A + 0.23 μ A	
	100 Hz to 5 kHz	0.35 mA/A + 0.23 μ A	
	(1 to 10) mA		
	(10 to 20) Hz	4.6 mA/A + 2.3 μ A	
	(20 to 45) Hz	1.7 mA/A + 2.3 μ A	
	(45 to 100) Hz	0.70 mA/A + 2.3 μ A	
	100 Hz to 5 kHz	0.35 mA/A + 2.3 μ A	
	(5 to 20) kHz	0.7 mA/A + 2.3 μ A	
	(20 to 50) kHz	4.6 mA/A + 4.6 μ A	
	(50 to 100) kHz	6.3 mA/A + 18 μ A	
	(10 to 100) mA		
	(10 to 20) Hz	4.6 mA/A + 23 μ A	
	(20 to 45) Hz	1.7 mA/A + 23 μ A	
	(45 to 100) Hz	0.7 mA/A + 23 μ A	
	100 Hz to 5 kHz	0.35 mA/A + 23 μ A	
	(5 to 20) kHz	0.70 mA/A + 23 μ A	
	(20 to 50) kHz	4.6 mA/A + 46 μ A	
	(50 to 100) kHz	6.4 mA/A + 0.17 mA	
	100 mA to 1 A		
	(10 to 20) Hz	4.6 mA/A + 0.23 mA	
(20 to 45) Hz	1.7 mA/A + 0.23 mA		
(45 to 100) Hz	0.7 mA/A + 0.23 mA		
100 Hz to 5 kHz	1.2 mA/A + 0.23 mA		
(5 to 20) kHz	3.5 mA/A + 0.23 mA		
(20 to 50) kHz	12 mA/A + 0.46 mA		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure ¹	10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ 10 MΩ 100 MΩ 1 GΩ	4.7 μΩ/Ω + 0.36 mΩ 11 μΩ/Ω + 0.69 mΩ 11 μΩ/Ω + 0.81 mΩ 12 μΩ/Ω + 5.8 mΩ 11 μΩ/Ω + 75 mΩ 17 μΩ/Ω + 2.3 Ω 13 μΩ/Ω + 0.17 kΩ 0.58 mΩ/Ω + 1.1 kΩ 5.8 mΩ/Ω + 12 kΩ	HP 3458A Precision Multimeter
Capacitance Measure ¹ 1 MHz	1 nF 10 nF 100 nF 1 μF	0.001 pF 0.84mF/F + 0.0002 pF 0.86 mF/F 8.6 mF/F	Keysight E4981A Capacitance Meter
Capacitance Measure 1 kHz	1 nF 10 nF 100 nf 1	0.82 mF/F + 0.000 04 nF 0.82 mF/F + 0.000 04 nF 0.82 mF/F + 0.000 04 nF 0.83 mF/F	Keysight E4981A Capacitance Meter
Dissipation Factor Measure ^{1,4}	0 DF	0.000 76	Keysight E4981A Capacitance Meter
Electrical Simulation of Thermocouple Instruments ¹	Type E (-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C Type J (-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1 200) °C Type K (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C (1 000 to 1 372) °C	0.59 °C 0.22 °C 0.2 °C 0.22 °C 0.27 °C 0.33 °C 0.22 °C 0.2 °C 0.23 °C 0.29 °C 0.4 °C 0.24 °C 0.22 °C 0.32 °C 0.48 °C	Fluke 5520A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Instruments ¹	Type R (0 to 250) °C	0.67 °C	Fluke 5520A Multifunction Calibrator
	(250 to 400) °C	0.42 °C	
	(400 to 1 000) °C	0.4 °C	
	(1 000 to 1 767) °C	0.48 °C	
	Type S (0 to 250) °C	0.56 °C	
	(250 to 1 000) °C	0.43 °C	
	(1 000 to 1 400) °C	0.44 °C	
	(1 400 to 1 767) °C	0.54 °C	
	Type T (-250 to -150) °C	0.74 °C	
	(-150 to 0) °C	0.3 °C	
	(0 to 120) °C	0.22 °C	
	(120 to 400) °C	0.2 °C	
	Type N (-200 to -100) °C	0.48 °C	
	(-100 to -25) °C	0.28 °C	
(-25 to 120) °C	0.25 °C		
(120 to 410) °C	0.24 °C		
(410 to 1 300) °C	0.33 °C		
Welding Power Supply ¹ Voltage Current	(0 to 50) V (0 to 500) A	0.02 V 4.4 A	Cannon Load Bank and Multimeter
DC Current Source – Clamp-on Ammeters	(20.5 to 1 000) A	3.3 mA/A + 0.46 A	Fluke 5520A Multifunction Calibrator and Coil
AC Current Source – Clamp-on Ammeters	(20.5 to 1 000) A (45 to 100) Hz	4.8 mA/A + 0.4 A	
	(20.5 to 600) A (100 to 1 000) Hz	5.3 mA/A + 0.27 A	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Micrometers ¹ Outside Inside	Up to 12 in	(79 + 13L) μin	Gage Blocks
	(1.5 to 12) in	(136 + 8.4L) μin	
Calipers ¹	Up to 40 in	(1 300 + 6L) μin	Gage Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Height Gages ¹	Up to 40 in	(90 + 10L) μin	Gage Blocks
Indicator Calibrator ¹	Up to 1 in	27 μin	Gage Blocks
Depth Gages ¹	Up to 24 in	(142 + 11L) μin	Gage Blocks
Length Standards ¹	Up to 28 in	(65 + 20L) μin	Gage Blocks and Analog Comparator
Coating Thickness Gauge ³	Up to 50 mils	0.21 mils	Supermicrometer and shims
Cylindrical Plugs	Up to 10 in	(5.3 + 17L) μin	Gage Blocks with ID/OD Comparator
Cylindrical Rings	(0.125 to 11) in	(0.8 + 17L) μin	
Indicators ¹	Up to 1 in	110 μin	Indicator Calibrator
Angle ¹	(0 to 90) °	0.013 °	10 in Sine Bar MET-SB-001
Roughness ¹	118 μin	5.2 μin	Roughness Standard

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Pressure – Hydraulic ¹	(2 to 10 000) psig	(0.084 + 0.000 17P) psi	Deadweight Tester
Pressure – Pneumatic ¹	(-14.7 to 3) psig (3 to 500) psig	0.008 8 psi (0.008 8 + 0.000 18P) psi	Pressure Module Deadweight Tester
Torque – Measure ¹	(2.5 to 25) lbf•in (25 to 250) lbf•in (25 to 250) lbf•ft (250 to 600) lbf•ft	0.33 % of reading 0.33 % of reading 0.62 % of reading 1.6 % of reading	Torque Calibrator MET-TW-001
Torque – Source ¹	(2.5 to 25) lbf•in (25 to 250) lbf•in (25 to 250) lbf•ft (250 to 600) lbf•ft	0.083 % of reading 0.1 % of reading 0.066 % of reading 0.14 % of reading	6 in Torque Wheel and arm Class F Weights MET-TW-002
Force ¹	Up to 250 lbf (250 to 1 000) lbf (1 001 to 10 000) lbf	(0.008 6 + 0.000 3F) lbf 0.29 lbf 18 lbf	Standard Weights, Load Cells



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Gas Flow ¹	(1 to 100) ccm (100 to 1000) ccm (1 to 250) lpm	(0.1 + 0.015X) ccm (5 + 0.01X) ccm (4.3 + 0.004 2X) lpm	Bubble Generator, Laminar Flow Element MET-AF-001
Hydraulic Flow ¹	(0.2 to 5) gpm	0.87 % of reading	Stopwatch/Prover MET-LF-001
Scales ¹ (0.0002 lbs resolution) (0.001 lbs resolution) (0.01 lbs resolution)	Up to 10 lb (6 to 60) lb (25 to 250) lb	(0.000 5 + 0.000 3F) lb (0.006 + 0.000 15F) lb (0.0002 + 0.013F) lb	Class 7 Weights
Analytical Balance ¹	Up to 100 g 100 g to 1 kg	1.3 mg 28 mg	Class 1 Weights
Hardness Testers	HRBW Low Middle High	1.2 HRBW 0.76 HRBW 0.72 HRBW	Indirect Verification using Hardness Blocks
	HREW Low Middle High	0.79 HREW 0.63 HREW 1.8 HREW	

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Temperature – Measure ¹	(-50 to 650) °C	0.04 °C	RTD, 1502 Indicator
	(650 to 1 350) °C	2.6 °C	Thermocouple, 5520A Multifunction Calibrator
Temperature – Source	(0 to 200) °C (200 to 500) °C	(0.04 + 0.000 06T) °C (0.07 + 0.000 02T) °C	RTD, 1502 Indicator Baths, Dry Well
Relative Humidity ¹	(20 to 100) % RH	1.2 %RH	Psychrometer

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Stopwatch ¹	1 s to 30 hr	(0.05 + 0.000 01S) s	GPS, 5334B Counter MET-SW-002
Frequency	10 MHz	1.2 x 10 ⁻¹² Hz	GPS Receiver
Optical Tachometer	(600 to 60 000) RPM	0.009 8 RPM + 3.3 μRPM/RPM	5520A Multifunction Calibrator

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

- Notes:
1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
 2. L = length in inches, P = pressure in psi, R = resolution of unit under test, S = time in unit of seconds, F = force in units of pounds, T = temperature in units degrees Celsius, c = flow rate in units scfm, l = flow rate in units slpm.
 3. 1 mil = 0.001 inch.
 4. Unitless linear measure.
 5. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1358.



Jason Stine, Vice President

